

Date of Certificate: 03/16/2017
Product Number: 5000-0020
Product Lot Number: 1196684
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1916026	03/13/2017	156	CS	20.7	26.0
1916989	03/14/2017	156	CS	20.1	26.0



Barry Srolis
Sr. Quality Manager