

Date of Certificate: 01/25/2017
Product Number: 5000-0020
Product Lot Number: 1193626
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1879738	01/17/2017	156	CS	19.9	25.3
1881399	01/23/2017	156	CS	20.0	25.5



Barry Srolis
Sr. Quality Manager