


---

Date of Certificate: 04/24/2017  
Product Number: 5000-0020  
Product Lot Number: 1199640  
Specified Dose Range: 19.0 - 28.0 KGY

| Run ID  | Receipt Date | Quantity | UOM | Delivered Min | Delivered Max |
|---------|--------------|----------|-----|---------------|---------------|
| 1939337 | 04/18/2017   | 192      | CS  | 20.9          | 26.4          |
| 1939721 | 04/19/2017   | 120      | CS  | 20.4          | 26.5          |



---

**Barry Srolis**  
Sr. Quality Manager