

Corning Incorporated - Life Sciences
271 County Route 64
Big Flats NY 14814 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name : 99445-16XX DCT 16 X 150MM	
Catalog Number : 99445-16XX	Manufacture Date : 2017-04-19
Lot ID : 10917446	
Expiration Date : Unconfirmed Validity	

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Alumino-Borosilicate, Type I, Class B, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Product Specification Standards - Product made in accordance with:
ASTM E890 - Standard Specification for Disposable Glass Culture Tubes

Animal Content - Product does not contain materials of animal origin.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Packaging Inspection - Pass
Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



David R Patrick
Quality Manager