

Corning Incorporated - Life Sciences  
271 County Route 64  
Big Flats NY 14814 USA  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b> : 7077 PIPET 2ML STERILE	<b>Manufacture Date</b> : 2017-05-23
<b>Catalog Number</b> : 7077-2N	
<b>Lot ID</b> : 14317699	
<b>Expiration Date</b> : Unconfirmed Validity	

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Alumino-Borosilicate, Type I, Class B, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Product Specification Standards** - Product made in accordance with:  
ASTM E714 - Standard Specification for Disposable Glass Serological Pipets

**Volumetric Tolerance** - Volumetric Accuracy at any graduation +/- 3 % of full capacity.

**Animal Content** - Product does not contain materials of animal origin.

**Sterility** - Product has been sterilized and parametrically released per the requirements of ANSI/AAMI/ISO 11135, "Sterilization of health care products - Ethylene oxide". Products meet a minimum Sterility Assurance Level (SAL) of  $10^{-6}$ .

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



David R Patrick  
Quality Manager