

Sorenson BioScience, Inc. - Life Sciences  
6507 South 400 West  
Salt Lake City UT 84107 USA  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information.

Page: 1 / 1

|                                               |                                      |
|-----------------------------------------------|--------------------------------------|
| <b>Product Name</b> : 3024 CYL 10ML           | <b>Manufacture Date</b> : 2017-07-06 |
| <b>Catalog Number</b> : 3024-10               |                                      |
| <b>Lot ID</b> : 18717005                      |                                      |
| <b>Expiration Date</b> : Unconfirmed Validity |                                      |

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Volumetric Tolerance** - Volumetric Accuracy at any graduation +/- 0.20 mL.

**Product Specification Standards** - Product made in accordance with:

ASTM E542 - Specification for Calibration of Laboratory Volumetric Apparatus

ASTM E694 - Standard Specification for Laboratory Glass Volumetric Apparatus, Class A Criteria

ASTM E1272 - Standard Specification for Laboratory Glass Graduated Cylinders

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass

Packaging Inspection - Pass

Dimensional Test - Pass

Capacity Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Miguel Flores  
Quality Manager