

Certificate of Quality

REF

454238

LOT

A17063CH

Product description: VACUETTE® TUBE 2 ml FX Sodium Fluoride / Potassium Oxalate
13x75 grey cap-white ring, non-ridged

Following inspections were carried out and passed:

visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed
chemical	<input checked="" type="checkbox"/>	passed

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

98/79/EC "Directive 98/79/EC of the European Parliament and of the Council of 27th Oct 1998 on in vitro diagnostic medical devices"

ISO 6710 "Single-use containers for venous blood specimen collection"

ISO 11137 "Sterilization of health care products - Radiation"

EN 14820 "Single-use containers for human venous blood specimen collection"

ISO 9001 "Quality management systems - Requirements"

ISO 13485 "Medical devices - Quality management systems - Requirements for regulatory purposes"



2018-10-11

STERILE

SAL 10⁻⁶

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2017-07-05

www.gbo.com