

Date of Certificate: 01/05/2018
Product Number: 5000-0020
Product Lot Number: 1215961
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2115852	01/03/2018	211	CS	19.9	25.5
2117848	01/03/2018	100	CS	19.9	26.8



Barry Srolis
Sr. Quality Manager