

Date of Certificate: 11/02/2017
Product Number: 5000-0020
Product Lot Number: 1209987
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2050114	09/26/2017	126	CS	20.4	25.6
2060274	10/11/2017	142	CS	20.1	26.7
2065465	10/17/2017	37	CS	19.6	26.1



Barry Srolis
Sr. Quality Manager