

Date of Certificate: 02/16/2018
Product Number: 5000-0020
Product Lot Number: 1220420
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2130228	01/23/2018	156	CS	20.6	26.0
2145349	02/13/2018	36	CS	21.2	26.7
2145199	02/14/2018	127	CS	20.3	26.1



Barry Srolis
Sr. Quality Manager