

Certificate Of Analysis

Whatman™

Product Information

Product Number: 28321-113

Product Name: 410 12.5CM 100/PK

Lot Number/Serial Number: 9851695

Date of Manufacture: 15-May-2017

Quality Control

| Test | Range | Results |
|------------------------|---------|---------|
| Grammage,gsm | 95-115 | 105 |
| Thickness, um/53kpa | 165-225 | 173 |
| Porosity, S/25ml/sq.in | 15-30 | 21 |

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2008 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AB



www.gelifesciences.com
GE Healthcare UK Limited
Amersham Place Little Chalfont
Buckinghamshire HP7 9NA UK