



# Certificate of Analysis

Becton Dickinson and Company  
 BD Diagnostic Systems  
 PO Box 999  
 Sparks MD 21152-0999 US

**Product Name** : Bottle Litmus Milk BBL 500G  
**Catalog Number** : 211343 **Manufacture Date**: 2018/01/25  
**Batch Number** : 8058939  
**Expiration Date** : 2021/11/30

- 01. Dehydrated Medium Powder Appearance: Fine, homogeneous and free of extraneous material
- 02. Solubility (100 g/L): Soluble in distilled or deionized water
- 03. Prepared Medium Appearance: Medium, purple gray; opaque
- 04. Biological Performance: Medium was prepared per label directions. Tubes were inoculated with the test organisms, incubated at 33-37°C for 7 days and gave cultural responses as indicated:

CULTURE	ATCC®	RESULT
Clostridium perfringens	13124	S, C, R
Lactobacillus acidophilus	4356	A, C

KEY: A=acid, pink  
 C=clot or curd  
 R=reduction, white  
 S=stormy fermentation  
 (strong ebullition of gas)

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C :		6.7	6.6	7.0
Bulk Lot Number :	-	7340827		

Animal source	Country of Origin	Tissue Category		
		BIC	SIC	ABC
Bovine	USA	IV	IV	MLK

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered. This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at <http://www.bd.com/regdocs>. For each batch of finished product that contains animal origin raw



# Certificate of Analysis

Becton Dickinson and Company  
BD Diagnostic Systems  
PO Box 999  
Sparks MD 21152-0999 US

Page: 2 of 2

**Product Name** : Bottle Litmus Milk BBL 500G  
**Catalog Number** : 211343 **Manufacture Date:** 2018/01/25  
**Batch Number** : 8058939  
**Expiration Date** : 2021/11/30

materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at <http://www.bd.com/regdocs> under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at <http://www.bd.com/regdocs> under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see [www.bd.com/ds/technicalCenter/regulatory.asp](http://www.bd.com/ds/technicalCenter/regulatory.asp).

Stability data for this product is on file.

The pH stated was obtained at the time of release, and the pH may vary within the stated range depending on the age of the product and the type of pH meter and probe used.

Charlotte Dannenfelser  
BD Life Sciences - Diagnostic Systems  
Quality Director, Microbiology  
Signature Date:2018/03/09