

Date of Certificate: 03/02/2018
Product Number: 5000-0020
Product Lot Number: 1219643
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2149833	02/20/2018	46	CS	20.5	26.5



Barry Srolis
Sr. Quality Manager