


---

Date of Certificate: 02/15/2018  
Product Number: 5000-1020  
Product Lot Number: 1219701  
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2146163	02/13/2018	156	CS	20.3	26.7
2145210	02/14/2018	156	CS	20.7	26.2



---

**Barry Srolis**  
Sr. Quality Manager