

Date of Certificate: 05/03/2018
Product Number: 5000-0020
Product Lot Number: 1219646
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2190079	04/24/2018	117	CS	19.9	25.1
2195027	05/01/2018	57	CS	20.8	25.9
2194926	05/01/2018	113	CS	21.1	26.2



Barry Srolis
Sr. Quality Manager