


---

Date of Certificate: 04/11/2018  
Product Number: 5000-0020  
Product Lot Number: 1226238  
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2180819	04/10/2018	156	CS	20.1	25.4
2179964	04/10/2018	156	CS	19.8	25.6



---

**Barry Srolis**  
Sr. Quality Manager