

Date of Certificate: 05/16/2018
Product Number: 5000-0020
Product Lot Number: 1229037
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2198329	05/09/2018	156	CS	19.5	25.4
2202867	05/15/2018	156	CS	21.1	25.6
2203630	05/16/2018	2	CS	19.9	24.6



Barry Srolis
Sr. Quality Manager