



Certificate of Analysis

1.00181.0002 Oxidase for the detection of cytochrome oxidase in microorganisms
Bactident®
Batch HC842219

| | Spec. Values | Batch Values |
|--|--------------|--------------|
| Cytochrome oxidase (Neisseria meningitidis ATCC 13090) | + | + |
| Cytochrome oxidase (Neisseria gonorrhoeae ATCC 9793) | + | + |
| Cytochrome oxidase (Pasteurella multocida ATCC 43137) | + | +(poor) |
| Cytochrome oxidase (Aeromonas hydrophila ATCC 7966 (WDCM 00063)) | + | + |
| Cytochrome oxidase (Pseudomonas alcaligenes ATCC 14909) | + | + |
| Cytochrome oxidase (Escherichia coli ATCC 11775 (WDCM 00090)) | - | - |
| Cytochrome oxidase (Staphylococcus aureus ATCC 25923 (WDCM 00034)) | - | - |
| Cytochrome oxidase (Citrobacter freundii ATCC 8090) | - | - |
| Cytochrome oxidase (Serratia marcescens ATCC 14756) | - | - |
| Cytochrome oxidase (Proteus mirabilis ATCC 29906 (WDCM 00023)) | - | - |

Date of release (DD.MM.YYYY) 21.06.2018

Expiry date (DD.MM.YYYY) 31.05.2019

Stefanie Fischer

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.