

Date of Certificate: 06/13/2018
Product Number: 5000-0020
Product Lot Number: 1226390
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2217949	06/05/2018	156	CS	20.0	25.9
2221682	06/13/2018	156	CS	19.9	26.0



Barry Srolis
Sr. Quality Manager