



Helping all people  
live healthy lives

Becton Dickinson Caribe LTD.  
BD Diagnostics Systems  
6 Vicks Drive Lot# 6  
Cayey PR 00737-2860 PR

### Certificate of Analysis

**SULFAMETHOXAZOLE+TRIMETHOPRIM**

**Susceptibility test discs**

Manufacturing Date: 11MAY18  
Catalog No. 231539/293595/296641

Lot No. **8107582**  
Expiration Date: 2021 05 31  
Release Date: 30/JUL/2018

This is to certify that representative samples of this lot of Sulfamethoxazole+ Trimethoprim were tested and gave satisfactory assay results established for this susceptibility disc by Becton Dickinson.

**Assay results:**

Agent	Code	Assay Range (mcg)	Avg. Potency	Uniformity (mm)
Sulfamethoxazole +Trimethoprim	SXT	SXT: 28.1 28.5 27.6 27.6 T.C.: 1.35 1.37	SXT:28 mcg T.C.:1.4 mcg	SXT: 1.0 1.0 0.4 0.4 T.C.: 1.7 1.3

BD Diagnostics Systems (BDDS) is an ISO 13485:2003 and ISO 9001:2000 Registered facility. BDDS products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by FDA's Quality System Regulations (QSR's). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release.

In addition, the batch record has been reviewed and approved which verifies that this batch was formulated and processed within the established specifications for this product and meet all BD Japan requirements. All Raw Data value meet Japan requirements.

Note: These discs were manufactured for Export / Stock.

Disposition: (Satisfactory/ Unsatisfactory): Satisfactory

Performed by: Wanda E. Amador Date: 31 JUL 18  
Approved by: [Signature] Date: 31 JUL 18

**STM A-S17, A-S17A**

31JUL18. doc  
weg