

Corning Incorporated - Life Sciences  
271 County Route 64  
Big Flats NY 14814 USA  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: 3160 DISH COMP 100X15	<b>Manufacture Date</b>	: 2018-02-28
<b>Catalog Number</b>	: 3160-101		
<b>Lot ID</b>	: 05918002		
<b>Expiration Date</b>	: Unconfirmed Validity		

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Animal Content** - Product does not contain materials of animal origin.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass  
Packaging Inspection - Pass  
Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Elizabeth M (Betty) Hansen  
Quality Manager