

Corning Incorporated - Life Sciences
271 County Route 64
Big Flats NY 14814 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name : 7077 PIPET 5ML STERILE	Manufacture Date : 2018-05-21
Catalog Number : 7077-5N	
Lot ID : 14118746	
Expiration Date : Unconfirmed Validity	

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Alumino-Borosilicate, Type I, Class B, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Product Specification Standards - Product made in accordance with:
ASTM E714 - Standard Specification for Disposable Glass Serological Pipets

Volumetric Tolerance - Volumetric Accuracy at any graduation +/- 3 % of full capacity.

Animal Content - Product does not contain materials of animal origin.

Sterility - Product has been sterilized and parametrically released per the requirements of ANSI/AAMI/ISO 11135, "Sterilization of health care products - Ethylene oxide". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-6} .

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:
Visual Inspection - Pass
Packaging Inspection - Pass
Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Elizabeth M (Betty) Hansen
Quality Manager