

Date of Certificate: 08/22/2018
Product Number: 5000-0020
Product Lot Number: 1234869
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2263144	08/14/2018	156	CS	20.3	25.7
2267225	08/22/2018	156	CS	19.8	25.4



Barry Srolis
Sr. Quality Manager