

Certificate of Analysis

Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Sparks MD 21152-0999 US Page: 1 of 3

Product Name : Bottle Fluid Thioglycollate Med 500G

Catalog Number : 225650 Manufacture Date: 2018/04/26

Batch Number : 8141537 Expiration Date : 2023/03/31

01. Dehydrated Medium Appearance: Light beige, free-flowing, homogeneous

02. Solubility: 2.98% solution, soluble in distilled or deionized water on boiling

03. Solution Appearance: Hot - light amber, clear. At room temperature - light amber, slightly opalescent, 10% or less of upper layer is medium pink. After shaking solution becomes pink throughout.

04. Cultural Response: Medium was prepared per label instructions. Tubes were inoculated with the test organisms and incubated at 30-35°C for 18-48 hours, or up to 76 hours if necessary.

TEST ORGANISMS ATCC® RECOVERY
Clostridium novyi 7659 good
Clostridium perfringens 13124 good
Staphylococcus aureus 25923 good

05. Meets USP/EP/JP performance specifications where applicable (harmonized chapters for "Sterility Tests.")

Inoculum of < 100 CFUs was used and tubes were incubated at 30-35°C for up to 3 days.

Pharmacopeia Growth Promotion:

Bacillus subtilis 6633 growth Bacteroides vulgatus 8482 growth Clostridium sporogenes 11437 growth Clostridium sporogenes 19404 growth Kocuria rhizophila 9341 growth Pseudomonas aeruginosa 9027 growth Staphylococcus aureus 6538 growth

06. Meets ISO11133 performance specifications. Tubes were inoculated with </= 100 CFUs and incubated at 37 ± 1°C for 18-24 hours.

ORGANISM ATCC® RESULT

Clostridium perfringens 13124 weak to good growth

07. Residual Solvents (CPMP/ICH/283/95): Typical Analysis for Fluid Thioglycollate medium did not detect any solvents.

Characteristic	Unit	Value	Lower Limit	Upper Limit	
Loss on Drying : pH at 25°C : Bulk Lot Number:	%	2.0 7.2 8106897	0.0 6.9	5.0 7.3	

Animal source	Country of	Tissue Category			
	Origin	BIC	SIC	ABC	

Creation Date: 2018/06/29 16:53:43



Certificate of Analysis

Becton Dickinson and Company BD Diagnostic Systems PO Box 999

Sparks MD 21152-0999 US Page: 2 of 3

Product Name : Bottle Fluid Thioglycollate Med 500G

Catalog Number : 225650 Manufacture Date: 2018/04/26

Batch Number : 8141537 Expiration Date : 2023/03/31

Avian China IV IV NDF Bovine New Zealand IV IV MLK Porcine USA III III IB

For an avian origin ingredient used in the manufacture of this product, the BD supplier is not able to confirm animal health. The supplier has confirmed that the ingredient undergoes extensive processing that includes exposure to strong acid, filtration, crystallization, and drying.

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered. This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at http://www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at http://www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at http://www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/ds/technicalCenter/regulatory.asp.

Creation Date: 2018/06/29 16:53:43



Certificate of Analysis

Page: 3 of 3

Becton Dickinson and Company BD Diagnostic Systems PO Box 999

Sparks MD 21152-0999 US

Manufacture Date: 2018/04/26

Expiration Date : 2023/03/31

Megan Richer

Megan Reeher

Senior Manager Quality Assurance

Creation Date: 2018/06/29 16:53:43