

Date of Certificate: 08/23/2018
Product Number: 5000-0020
Product Lot Number: 1234876
Specified Dose Range: 19.0 - 28.0 KGY

| Run ID | Receipt Date | Quantity | UOM | Delivered Min | Delivered Max |
|---------|--------------|----------|-----|---------------|---------------|
| 2266503 | 08/21/2018 | 156 | CS | 19.6 | 25.0 |
| 2267242 | 08/22/2018 | 156 | CS | 19.4 | 25.6 |



Barry Srolis
Sr. Quality Manager