



Quality Certificate

VWR® STERILE SAMPLING BAGS

As per the manufacturer, the below product meets the following criteria:

North American Catalog No:	470236-300
Manufacture Lot Number:	101117
Description:	Wards sample bag pe 178 x 76mm 4oz pk 50
Date of Manufacture:	October 11, 2017
Sterility Expiration:	October 11, 2022
Expiration Date:	October 11, 2022

Product Specifications

Material	Polyethylene, stainless steel wire. Heavy metals content is compliant with EC/94/62	APPROVED
Appearance	Translucent, no visible impurities.	APPROVED
Food Contact	Raw material is compliant with EC N°1935/2005	APPROVED
Measurement	Length: 178mm; Width: 76mm	APPROVED
Temperature Resistance	Up to 100°C maximum	APPROVED
Performance	Holds 120mL	APPROVED
Sterilization	Gamma Irradiation, SAL 10 ⁻³	APPROVED

Quality System Compliance

Products are manufactured under the ISO 13485 system with SOPs and class 100,000 clean-room quality guarantee. Stringent quality control protocols are followed at every process of production ranging from qualitative raw materials and process of production through final products.

Sterilization

The products have been sterilized with Ethylene Oxide and found to conform to the sterility requirements of EN ISO 11135-1:2006.

QC Testing

Products are inspected and controlled through the entire production process in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed by qualified personnel for product release.

Signed:

Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: October 13, 2017