

Date of Certificate: 10/24/2018
Product Number: 5000-0020
Product Lot Number: 1239116
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2308931	10/24/2018	156	CS	20.3	26.1
2308830	10/24/2018	156	CS	20.6	27.0



Barry Srolis
Sr. Quality Manager