

Date of Certificate: 11/21/2018
Product Number: 5000-0020
Product Lot Number: 1241070
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2327060	11/21/2018	312	CS	20.2	26.3



Barry Srolis
Sr. Quality Manager