

Certificate Of Analysis

Whatman™

Product Information

Product Number: 1001-185

Product Name: GR 1 18.5CM 100/PK

Lot Number/Serial Number: 16915348

Date of Manufacture: 17-Oct-2018

Quality Control

Test	Range	Results
Grammage,gsm	78-96	87
Thickness, um/53kpa	154-206	172
Porosity, S/200ml/sq.in	15-37	31

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AC



imagination at work

www.gelifesciences.com

GE Healthcare UK Limited

Amersham Place Little Chalfont

Buckinghamshire HP7 9NA UK