

Certificate Of Analysis

Whatman™

Product Information

Product Number: 3001-604

Product Name: GR 1CHR 1CMx100M

Lot Number/Certificate No: 969350301

Parent Reel Job No: 9676324

Manufacture Date of Parent Reel: 15 SEP 2015

Reel Number: RL1523317

Quality Control

| Test | Range | Results | | |
|-------------------------|---------|---------|-----|-----|
| | | Mean | Min | Max |
| Grammage,gsm | 80-93 | 87 | 85 | 88 |
| Thickness, µm@53Kpa | 160-196 | 177 | 169 | 181 |
| Porosity,s/200ml/sq" in | 13-31 | 23 | 19 | 31 |

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to **BS EN ISO9001:2008** and tested in accordance with documented quality procedures and approved as a result of meeting the required specification

Electronic signature

This document has been electronically produced and is valid without a signature.

Version AC